Date: 22/01/2024

CPG1500 0...700 bar gauge G 1/2 B

CPG1500 0...700 bar gauge G 1/2 B

Item number: 46885349

Quantity: 1 pcs.

Order code: CPG1500-ST-Z-S-BG570-GDTZ-53-CLHZ-Z-ZZZZ Z W
CERT-PR-DE1-AP03C1C-Z-C12-ZZ-Z 1

Overview:

CPG1500 0...700 bar gauge G 1/2 B
Precision digital pressure gauge - Model CPG1500
Specifications according to data sheet: CT 10.51
Explosion protection: without
Measuring range: 0...700 bar gauge
Process connection: G 1/2 B
Wetted parts: stainless steel, Elgiloy Sensor
Accuracy: 0.05 % FS
Type of certificate: Acceptance test certificate 3.1 in accordance with DIN EN 10204
Communication: Bluetooth®
Info Bluetooth®: Please bear in mind the relevant radio approvals in your country.
Integrated data logger
incl. protective housing cap

Calibration of pressure measuring instruments

Type of certificate: Acceptance test certificate 3.1 in accordance with DIN EN 10204
Type of device under test: Electrical measuring device with display (digital pressure gauge, pressure controller, etc.)
Accuracy: > 0,02 % … < 0,1 %
Numbers of measurement series: 2
Number of measuring points per measurement series: 9
Pressure type - pressure: Relative pressure (over- or underpressure)
Calibration range - pressure: > 400 bar … <= 1.600 bar

Documentation: Calibration (documentation of the values)
1. Language of certificate: DE
2. Language of certificate: GB

Details:

CPG1500 0...700 bar gauge G 1/2 B
Precision digital pressure gauge - Model CPG1500
Specifications according to data sheet: CT 10.51
Explosion protection: without
Measuring range: 0...700 bar gauge
Process connection: G 1/2 B
Wetted parts: stainless steel, Elgiloy Sensor
Accuracy: 0.05 % FS

© 2024 - WIKA Instruments s.a.r.l.

Type of certificate: Acceptance test certificate 3.1 in accordance with DIN EN 10204
Communication: Bluetooth®
Info Bluetooth®: Please bear in mind the relevant radio approvals in your country.
Integrated data logger
incl. protective housing cap

Calibration of pressure measuring instruments

Type of certificate: Acceptance test certificate 3.1 in accordance with DIN EN 10204
Type of device under test: Electrical measuring device with display (digital pressure gauge, pressure controller, etc.)
Accuracy: > 0,02 % … < 0,1 %
Numbers of measurement series: 2
Number of measuring points per measurement series: 9
Pressure type - pressure: Relative pressure (over- or underpressure)
Calibration range - pressure: > 400 bar … <= 1.600 bar

Documentation: Calibration (documentation of the values)
1. Language of certificate: DE
2. Language of certificate: GB

© 2024 - WIKA Instruments s.a.r.l.