*Date: 24/04/2023*

# Adapter system for measuring instruments with G 1 hygienic connection; Model 910.61

Adapter system for measuring instruments with G 1 hygienic connection; Model 910.61

| | |
|---|---|
| Item number: | 48794369 |
| Quantity: | 1 pcs. |
| Order code: | 910.61-CP115D4C2Z3<br>LABEL-PG1DE-ZZZZZ- ZZZZZ<br>CERT31-EF4A-BB- - - - - - - - - - - |
| Overview: | Adapter system for measuring instruments with G 1 hygienic connection; Model 910.61<br>Specifications according to data sheet: AC 09.20<br>Clamp connection - Tri-clamp<br>Model 910.61 1 1/2" PN 40 pipe standard: Pipe dimensions per BS 4825 Part 3 or O.D.-Tube<br>Upper body material / main body material: Stainless steel 1.4435 (316L); UNS S31603<br><br>Inspection certificate acc. to DIN EN 10204 3.1<br>Certificate Language: English and French<br>Certificate for: Diaphragm seal<br>Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis<br><br>Label: WIKA standard without / with Tag-number |
| Details: | Adapter system for measuring instruments with G 1 hygienic connection; Model 910.61<br>Specifications according to data sheet: AC 09.20<br>Diaphragm seal: Model 910.61<br>Process connection type: Clamp connection<br>Process connection specification: Tri-clamp<br>Process connection: 1 1/2"<br>Nominal pressure rating: PN 40<br>pipe standard: Pipe dimensions per BS 4825 Part 3 or O.D.-Tube<br>Upper body material / main body material: Stainless steel 1.4435 (316L); UNS S31603<br>Inspection certificate acc. to DIN EN 10204 3.1<br>Certificate Language: English and French<br>Certificate for: Diaphragm seal<br>Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis |

*© 2023 - WIKA Instruments s.a.r.l.*

*Date: 24/04/2023*

Label: WIKA standard without / with Tag-number

*© 2023 - WIKA Instruments s.a.r.l.*