Date: 25/04/2023

233.30.100 60 bar G1/2B lower mount

Bourdon Tube Pressure Gauges, Safety Pattern

Item number: 48807218

Quantity: 1 pcs.

Order code: 233.30-E-BG460Z-GD-UZHZCZZ-ZZZ1ZZZ 213D3AZIYG0 Z2Z-IZZZZZZ12ZZZ
CERT31-EF2D-AB- - - - - - - - - - -
LABEL-PG1DE-ZZZZZ

Overview:
233.30.100 60 bar G1/2B lower mount
Measuring System 316L
Ingress protection: IP 65
class 1.0
Window: laminated safety glass
Case filling fluid: glycerine
Permissible medium temperature: + 100 °C
Permissible ambient temperature: -20 … +60 °C
Pressure Gauge Standard: International (standard Europe)
WIKA
with Serial number
Inspection certificate 3.1 per EN 10204
Certificate type: Inspection certificate acc. to DIN EN 10204 3.1
Certificate Language: English and French
Certificate for: pressure gauge
Accuracy: Calibration values listed
Label: WIKA standard without / with Tag-number

Details:
Bourdon Tube Pressure Gauges, Safety Pattern
Series EN 837-1
Model: 233.30
Specifications according to data sheet: PM 02.04
Measuring system: stainless steel
Case filling: with
Nominal size: 100 mm
Measuring range: gauge pressure range
Scale range: 0...60 bar
Process connection: G 1/2 B
Connector location: lower mount
Window: laminated safety glass
Pointer: standard pointer
Accuracy class: class 1.0
Overpressure safe: standard
Case material: Stainless steel 1.4301
Case surface: bare
Material pressure element: Measuring System 316L

© 2023 - WIKA Instruments s.a.r.l.

Case filling fluid: glycerine
Ingress protection: IP 65
Permissible medium temperature: + 100 °C
Permissible ambient temperature: -20 … +60 °C
Movement: stainless steel
Packaging: single packing
Pressure Gauge Standard: International (standard Europe)
Manufacturer logo: *WIKA*
with Serial number
Permissible ambient temperature: -20 … +60 °C
Inspection certificate 3.1 per EN 10204
Certificate type: Inspection certificate acc. to DIN EN 10204 3.1
Certificate Language: English and French
Certificate for: pressure gauge
Accuracy: Calibration values listed

© 2023 - WIKA Instruments s.a.r.l.