TPS2000 Series Oscilloscope
P2220 Probe Safety

Warning Sheet

063-3804-00



WARNING: *Do not float the P2220 probe reference lead to >30 V_{RMS}. Use the P5120 probe (floatable to 600 V_{RMS} CAT II or 300 V_{RMS} CAT III), or a similarly rated, passive, high voltage probe, or an appropriately rated, high voltage, differential probe when floating the reference lead above 30 V_{RMS}, subject to the ratings of such high voltage probe.*

063380400