Unigloves

Product specification sheet

Revision 5 | 24 October 2017

1. Product code: GA005*

2. Product description: PRO.TECT Orange HD

3. Material: Nitrile

4. Inspection parameters

Criteria	Inspection level	AQL	Standard
Dimensions	S-2	4	EN455-2
Physical properties	S-2	4	EN455-2
Water tight test (1,000ml)	G-1	1.5	EN455-1
Visual defects	G-1	2.5	In-house

5. Glove dimensions

Dimensions	Size (mm)				
	XS	S	M	L	XL
Product code	-	-	GA0053	GA0054	GA0055
Palm width	-	-	95 ± 10	110 ± 10	≥ 110
Length	-	-	≥ 240	≥ 240	≥ 240

6. Glove thickness

Point of thickness measurement	Nominal single wall (mm)
Cuff (at beaded end)	0.09
Palm (at centre of palm)	0.13 (flat area) 0.22 (raised area)
Fingertip (at fingertip)	0.15

7. Physical properties

Parameters	Before aging	After aging
Median force at break (N)	Min 6.0	Min 6.0
Tensile strength (MPa)	Min 14.0	Min 14.0
Ultimate elongation (%)	Min 500	Min 400

8. Proteins: None

9. Powder residue: Maximum 2.0 mg / glove

10. Cuff type: Beaded

11. Design: Ambidextrous

12. Surface: Diamond Textured

13. Colour: Orange

14. Quality assurance: Manufactured under ISO 9001:2008 and 13485:2012

The above data represents the product norm, which may be changed without notice and is intended as a guide only.