Date: 24/04/2023

232.30.063 4 bar G1/4B lower mount

Bourdon Tube Pressure Gauges, Safety Pattern

Item number: 48793157

Quantity: 1 pcs.

Order code: 232.30-C-BG340Z-GB-UZHZCZZ-ZZZ1ZZZ 313D41ZK3W0 Z2Z-IZZZZZZ121ZZ
CERT22-EF2D-FB- - - - - - - - - - -
CERT31-EF2D-BB- - - - - - - - - - -
LABEL-PG1DE-ZZZZZ

Overview:
232.30.063 4 bar G1/4B lower mount
Measuring System 316 SS
Ingress protection: IP 66
class 1.6
Window: laminated safety glass
Permissible medium temperature: + 200 °C
Permissible ambient temperature: -40 … +60 °C
Pressure Gauge Standard: International (standard Europe)
stop-pin at zero
WIKA
with Serial number
Test report 2.2 per EN 10204
Certificate Language: English and French
Certificate for: pressure gauge
Ingress protection: Ingress protection IP66
Inspection certificate 3.1 per EN 10204
Certificate type: Inspection certificate acc. to DIN EN 10204 3.1
Certificate Language: English and French
Certificate for: pressure gauge
Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis
Label: WIKA standard without / with Tag-number

Details:
Bourdon Tube Pressure Gauges, Safety Pattern
Series EN 837-1
Model: 232.30
Specifications according to data sheet: PM 02.04
Measuring system: stainless steel
Case filling: without
Nominal size: 063 mm
Measuring range: gauge pressure range
Scale range: 0...4 bar
Process connection: G 1/4 B
Connector location: lower mount
Window: laminated safety glass

© 2023 - WIKA Instruments s.a.r.l.

Pointer: standard pointer
Accuracy class: class 1.6
Overpressure safe: standard
Case material: Stainless steel 1.4301
Case surface: bare
Material pressure element: Measuring System 316 SS
Case filling fluid: prepared for filling
Ingress protection: IP 66
Permissible medium temperature: + 200 °C
Permissible ambient temperature: -40 … +60 °C
Movement: stainless steel
Packaging: single packing
Pressure Gauge Standard: International (standard Europe)
Manufacturer logo: *WIKA*
stop pin on dial: at zero
with Serial number
Permissible ambient temperature: -40 … +60 °C
Test report 2.2 per EN 10204
Certificate Language: English and French
Certificate for: pressure gauge
Ingress protection: Ingress protection IP66
Inspection certificate 3.1 per EN 10204
Certificate type: Inspection certificate acc. to DIN EN 10204 3.1
Certificate Language: English and French
Certificate for: pressure gauge
Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis

© 2023 - WIKA Instruments s.a.r.l.