SA-11; 0 ... 6 bar; 4...20 mA, 2-W

Pressure transmitter for sanitary applications

Item number: 48667815

Quantity: 1 pcs.

Order code: SA-11-A-BBH-SJ-AGM4ZZ-1ZZ W101Z-F2B-GSSAK6
CERT31-EG1P-BB- - - - - - - - - - - SA
LABEL-PG1DE-ZZZZZ- ZZZZZ

Overview:
SA-11; 0 ... 6 bar; 4...20 mA, 2-W
@ 10 ... 30 V DC; Accuracy: 0.5% of span
union nut DIN 11851 with conical coupling DN 25;
circular connector M12x1, 4-pin UB=1, 0V=3
Specifications according to data sheet: PE 81.80
System fill fluid: KN 77 Synthetic oil
Inspection certificate 3.1 per EN 10204
Certificate type: Inspection certificate acc. to DIN EN 10204 3.1
Certificate Language: English and German
Certificate for: pressure transmitter
Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis

Label: WIKA standard without / with Tag-number

Details:
Pressure transmitter for sanitary applications
Model SA-11
Specifications according to data sheet: PE 81.80
Wetted parts: stainless steel 1.4435
Case: stainless steel
Permissible medium temperature: -20 ... +150 °C (-4 ... +302 °F)
Output signal: 4 ... 20 mA, 2-wire
Unit: bar
Measuring range: 0 ... 6 bar
Process connection: union nut DIN 11851 with conical coupling DN 25
Accuracy: 0.5% of span
Electrical connection: circular connector M12x1, 4-pin
Certificates: quality certificates
Power supply: 10 ... 30 V DC
Pin assignment: UB=1, 0V=3
Made in Germany on label
WIKA standard - specification according to data sheet
System fill fluid: KN 77 Synthetic oil
Import documentation CIS-States: without
Inspection certificate 3.1 per EN 10204
Certificate type: Inspection certificate acc. to DIN EN 10204 3.1
Certificate Language: English and German

© 2023 - WIKA Instruments s.a.r.l.

Date: 11/04/2023

Certificate for: pressure transmitter
Material certificate: Wetted parts material (metal contents) with mill certificate incl. chemical analysis

Label: WIKA standard without / with Tag-number

© 2023 - WIKA Instruments s.a.r.l.