Date: 24/04/2023

233.50.063 10 bar Diaphragm Seal lower mount

Bourdon Tube Pressure Gauges, Stainless Steel

Item number: 48663363

Quantity: 1 pcs.

Order code: DSS1-MS990.22-123X50-DZZZD-C7S65ZK5-BZAZZZZZ WAZ-Z
LABEL-PG1DE-ZZZZZ- ZZZZZ
233.50-C-BG410Z-CA-UZEZAZZ-ZZZZZZZ 313D4AZIXX0 Z2Z-IZZZZZZ121ZZ
H990.22-TB110C2C2AZ505ZZZ Z
CERT31-EG41-BA- - - - - - - - - - -
P1-E-PA

Overview: 233.50.063 10 bar Diaphragm Seal lower mount
Measuring System 316 SS
Ingress protection: IP 65
class 1.6
Window: polycarbonate
adjustable pointer
Case filling fluid: glycerine
Pressure Gauge Standard: International (standard Europe)
stop-pin at zero *WIKA*
with Serial number

Diaphragm seal system with 1 diaphragm seal
Assembly specification High-Side: Direct assembly
System fill fluid: KN 59 Neobee M-20; for food and pharmaceutical applications;FDA
Process temperature min.: + 10°C
Process temperature max.: + 80°C; for sterilisation (SIP) max. + 130°C
Ambient temperature pressure meas. min: + 10°C
Ambient temperature pressure meas. max: + 40°C

Diaphragm seal for hygienic applications
Tri-Clamp following BS 4825 Part 3
Model 990.22 DN 25.4 (pipe dimensions 25.4 x 1.6)
pipe standard: Pipe dimensions per BS 4825 part 3 or O.D.-Tube
Upper body material / main body material: Stainless steel 1.4435 (316L); UNS S31603
Diaphragm material: Stainless steel 1.4435 (316L); UNS S31603
Surface roughness wetted parts, welding seam excluded: Ra ≤ 0,76 µm (Ra 0.30 µin.) according to ASME BPE SF3
Clamp diameter: 50.5 mm

Calibration Test Report

© 2023 - WIKA Instruments s.a.r.l.

Language: English and German
3 test points: Begin/middle/end of range
Certificate type: Inspection certificate acc. to DIN EN 10204 3.1
Certificate Language: English and German
Certificate for: Diaphragm seal system with pressure measuring instrument
Material certificate: Wetted parts material (metal contents)

Label: WIKA standard without / with Tag-number

Details:

Bourdon Tube Pressure Gauges, Stainless Steel
Series
Model: 233.50
Specifications according to data sheet: PM 02.02
Measuring system: stainless steel
Case filling: with
Nominal size: 063 mm
Scale range: 0...10 bar
Process connection: Diaphragm Seal
Connector location: lower mount
Window: polycarbonate
Pointer: adjustable pointer
Accuracy class: class 1.6
Overpressure safe: standard
Case material: Stainless steel 1.4301
Case surface: bare
Material pressure element: Measuring System 316 SS
Case filling fluid: glycerine
Ingress protection: IP 65
Movement: stainless steel
Packaging: single packing
Pressure Gauge Standard: International (standard Europe)
Manufacturer logo: *WIKA*
with Serial number
stop pin on dial: at zero
Diaphragm seal system with 1 diaphragm seal
Assembly specification High-Side: Direct assembly
System fill fluid: KN 59 Neobee M-20; for food and pharmaceutical applications;FDA
Process temperature min.: + 10°C
Process temperature max.: + 80°C; for sterilisation (SIP) max. + 130°C
Ambient temperature pressure meas. min: + 10°C
Ambient temperature pressure meas. max: + 40°C
Diaphragm seal for hygienic applications
Specifications according to data sheet: DS 99.41
Diaphragm seal: Model 990.22
Process connection type: Tri-Clamp following BS 4825 Part 3
Process connection: DN 25.4 (pipe dimensions 25.4 x 1.6)

© 2023 - WIKA Instruments s.a.r.l.

pipe standard: Pipe dimensions per BS 4825 part 3 or O.D.-Tube
Upper body material / main body material: Stainless steel 1.4435 (316L); UNS S31603
Diaphragm material: Stainless steel 1.4435 (316L); UNS S31603
Surface roughness wetted parts, welding seam excluded: Ra ≤ 0,76 µm (Ra 0.30 µin.) according to ASME BPE SF3
Clamp diameter: 50.5 mm
Test-Report
Test-Report: Calibration Test Report
Language: English and German
Type of test: 3 test points: Begin/middle/end of range
Certificate type: Inspection certificate acc. to DIN EN 10204 3.1
Certificate Language: English and German
Certificate for: Diaphragm seal system with pressure measuring instrument
Material certificate: Wetted parts material (metal contents)

Label: WIKA standard without / with Tag-number

© 2023 - WIKA Instruments s.a.r.l.