Date: 22/04/2023

733.14.160 6 bar 2 x G 1/2 female

Differential pressure gauge, multi purpose, with diaphragm element,

Item number:	48798736
Quantity:	1 pcs.
Order code:	733.14-AHAF-HBG360Z-AD-UZZZASZZ-1ZZZZ 7B 3G3DIZ-IZZZZZZ112ZZ CERT31-EF2D-CE- - - - - - - - - - - LABEL-PG1DE-ZZZZZ
Overview:	733.14.160 6 bar 2 x G 1/2 female lower mount Explosion protection: ATEX Ex h Gas Zone 1 / Dust Zone 21 (II 2G Ex h IIC T6 ... T1 Gb X / II 2D Ex h IIIC T85 °C … T450 °C Db X) PN 100 bar Wetted parts: stainless-steel or NiCr-alloy (Inconel) & FPM with external zero adjustment Ingress protection: IP 65 class 1.6 glycerine-water-mixture Permissible ambient temperature: -20...+60 °C Permissible medium temperature: -20...+100 °C Pressure Gauge Standard: International (standard Europe) stop-pin at 6 o'clock *WIKA* Serial number *...* Inspection certificate 3.1 per EN 10204 Certificate type: Inspection certificate acc. to DIN EN 10204 3.1 Certificate Language: English and French Certificate for: pressure gauge Pressure and stability test: max. static (working) pressure; duration 15 min Label: WIKA standard without / with Tag-number
Details:	Differential pressure gauge, multi purpose, with diaphragm element, high overpressure limit Model: 733.14 Specifications according to data sheet: PM 07.13 Case: stainless steel Bezel ring: stainless steel Movement: stainless steel Explosion protection: ATEX Zone: Gas Zone 1 / Dust Zone 21 (II 2G Ex h IIC T6 ... T1 Gb X / II 2D Ex h IIIC T85 °C … T450 °C Db X) Ignition protection type: Ex h Wetted parts: stainless-steel or NiCr-alloy (Inconel) & FPM Case filling: with Nominal size: 160 mm

© 2023 - WIKA Instruments s.a.r.l.

Overpressure safety/working pressure max.: PN 100 bar
Scale range: 0...6 bar
Process connection: 2 x G 1/2 female
Connector location: lower mount
Permissible ambient temperature: -20...+60 °C
Permissible medium temperature: -20...+100 °C
Accuracy class: class 1.6
Case filling fluid: glycerine-water-mixture
Case surface: bare
Ingress protection: IP 65
Pressure Gauge Standard: International (standard Europe)
Manufacturer logo: *WIKA*
serial number on dial: Serial number *...*
stop pin on dial: at 6 o'clock
with external zero adjustment
Inspection certificate 3.1 per EN 10204
Certificate type: Inspection certificate acc. to DIN EN 10204 3.1
Certificate Language: English and French
Certificate for: pressure gauge
Pressure and stability test: max. static (working) pressure; duration 15 min

© 2023 - WIKA Instruments s.a.r.l.